

CR NO . 336218



log no - 2588 جی

88 شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاکس : ۲۴۵۶۷۷۹۲ / ۲۴۷۴۳۰۱۹ (+۹۶۵)

ص.ب: ٢٣٥٥ الجبراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ۴۳۶۲۱۸

No.: 58284

Customer <u>Alorif Hospital.</u>		Equipment <u>CSSD washer</u>	
		Model No. <u>Uniclean PL11 - 15</u>	
		Serial No. <u>230506</u>	
Authorized Person		Manufacturer <u>MMM</u>	
Customer Complaint		Location <u>CSSD</u>	
<u>Error 91 - Temperature limiter heating.</u>		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐ _____		☐ Installation ☐ Warranty	
		☒ Repair ☐ On Charge	
		☐ Maintenance ☐ Contract	
		☐ Upgrade ☒ <u>EM</u>	
		Time Spent	
Remarks <u>Service call attended.</u> <u>while we inspect the unit the error 91 occurs.</u> <u>we checked the unit and found that the heating</u> <u>element issue. we reset the heating element</u> <u>and tried. now the unit is working.</u> <u>checked and confirmed that the unit is</u> <u>working properly.</u>			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer <u>Nixon</u> ATCO Technology Co. Signature <u>[Signature]</u> Date <u>10/02/2024</u>		Authorized Person <u>[Signature]</u> Signature _____ Date _____	
		Customer / User Signature _____ Date _____	