

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63838

Customer	Alor Hospital			Equipment	Washer		
				Model No.	Uniclean PL11		
				Serial No.	230506		
Authorized Person				Manufacturer	mmm		
Customer Complaint				Location	CSSB		
				Contract / Order No.			
Services / Check Carried out				☐ Installation		☐ Warranty	
				☒ Repair		☐ On Charge	
☐ Inspection / Cleaning				☐ Optics Checks		☐ Maintenance	
☐ Calibration / Performance				☐ Pneumatic Checks		☐ Contract	
☐ Interface / Communication				☐ Hydraulic Checks		☐ Upgrade	
☐ Safety Checks				☐ Operational Error		☐ _____	
☐ Electrical / Electronics				☐ Mechanical Checks			
☐ Accessories Check				☐ _____			
				Time Spent			
Remarks							
Service call attended, we inspect the unit and found that the unloading door is not opening. we manually opened the door and reset the relay. now the unit is working properly. unit under observation.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer		Authorized Person			Customer / User		
Mixon							
ATCO Technology Co.							
Signature		Signature			Signature		
Date		Date			Date		
25/02/2026		17/2/25					