

# ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

## SERVICE / INSTALLATION REPORT

No.: 64047

|                    |               |                      |                 |
|--------------------|---------------|----------------------|-----------------|
| Customer           | Alor Hospital | Equipment            | Autoclave       |
|                    |               | Model No.            | Select mat PL   |
|                    |               | Serial No.           | 220824 & 220823 |
| Authorized Person  |               | Manufacturer         | mmm             |
| Customer Complaint |               | Location             | CSSD            |
| Stain on Packing   |               | Contract / Order No. |                 |

|                                                 |                                         |                                 |
|-------------------------------------------------|-----------------------------------------|---------------------------------|
| Services / Check Carried out                    | <input type="radio"/> Installation      | <input type="radio"/> Warranty  |
|                                                 | <input checked="" type="radio"/> Repair | <input type="radio"/> On Charge |
| <input type="radio"/> Inspection / Cleaning     | <input type="radio"/> Maintenance       | <input type="radio"/> Contract  |
| <input type="radio"/> Calibration / Performance | <input type="radio"/> Upgrade           | <input type="radio"/> _____     |
| <input type="radio"/> Interface / Communication | Time Spent                              |                                 |
| <input type="radio"/> Safety Checks             |                                         |                                 |
| <input type="radio"/> Electrical / Electronics  |                                         |                                 |
| <input type="radio"/> Accessories Check         |                                         |                                 |
| <input type="radio"/> Optics Checks             |                                         |                                 |
| <input type="radio"/> Pneumatic Checks          |                                         |                                 |
| <input type="radio"/> Hydraulic Checks          |                                         |                                 |
| <input type="radio"/> Operational Error         |                                         |                                 |
| <input type="radio"/> Mechanical Checks         |                                         |                                 |

Remarks we inspect the unit and cleaned all pipings. cleaned the chamber and steam inlet. we performed test cycle and the unit under observation.

| No. | Spare Parts Used   | Part No. | Qty. |
|-----|--------------------|----------|------|
|     | still small amount |          |      |
|     | amount spot        |          |      |
|     | noted              |          |      |

|                     |                   |                 |
|---------------------|-------------------|-----------------|
| Service Engineer    | Authorized Person | Customer / User |
| Nidhan/ Jinslamun   |                   |                 |
| ATCO Technology Co. |                   |                 |
| Signature           | Signature         | Signature       |
| Date 03/02/2025     | Date 18/2/25      | Date            |