

ATCO Technology Co.

Tel: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة اتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58675

Customer	Alor Hospital Jahra	Equipment	Mobile X Ray Unit		
		Model No.	Cyber mobil TSLD		
		Serial No.	T 7296		
Authorized Person		Manufacturer	Primax		
Customer Complaint	Collimator issue	Location	XRay		
		Contract / Order No.			
Services / Check Carried out		☐ Installation	☐ Warranty		
		☒ Repair	☒ On Charge		
		☐ Maintenance	☐ Contract		
		☐ Upgrade	☐ _____		
		Time Spent			
Remarks					
Service call attended. Checked the unit and the collimator plate inside. It was removed moved condition. Fixed the plate properly. Checked and found that the unit is working properly.					
No.	Spare Parts Used	Part No.	Qty.		
Service Engineer		Authorized Person			
Signature		Signature			
Date		Date			
Customer / User		Signature			
		Date			