

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61080

Customer	Al orf Hospitel	Equipment	Auto clare
		Model No.	sterile clare
		Serial No.	18X25721
Authorized Person		Manufacturer	XXXXXX Cominox
Customer Complaint	The machine stop working during the cycle.	Location	Dental
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
		Time Spent	

Remarks Service call attended.
checked the Pices and there was one damaged and stuck inside
the machine, we change it and we checked the cable power and
found the Pices also damaged, so we changed.
checked the machine working well. Done Boire and cycle.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Moussa / Nixon ATCO Technology Co. Signature Date 22-08-2024	Authorized Person Signature Date	Customer / User د. عبد الرحمن السعيد مستشفى مسجل اول الاستشارة الصناعية 22019100056 Signature Date
---	--	---