

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2591

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61603

Customer	Aloof Hospital	Equipment	Autoclave
		Model No.	Selectom 4 PL
		Serial No.	220823
Authorized Person		Manufacturer	MMM
Customer Complaint		Location	CSSD

Services / Check Carried out	☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check	☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐	Contract / Order No.	
			☐ Installation ☒ Repair ☐ Maintenance ☐ Upgrade	☐ Warranty ☐ On Charge ☐ Contract ☐
Time Spent				

Remarks	Service call attended we inspect the unit and cleared the chamber we drained out the Steam Generator water and refill the new water. The occurrence is due to the water quality. we kept the unit under observation.
---------	---

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Mizor/Mousc ATCO Technology Co. Signature Date 11/09/2024	Authorized Person Signature Date	Customer / User Signature Date
--	--	--------------------------------------