

شركة وربة للتجهيزات الطبية ذ.م.م

WARBA MEDICAL SUPPLIES CO. W.L.L.

تلفون المكتب: ٢٢٣٢٣٨٥٦ / ٢٢٣٢٣٨٥٧ - فاكس: ٢٢٣٢٣٨٦٤ - س.ت: ٢٧٤٦٨ - ص.ب: ٢٦٢٦٧ الصفاة ١٣١٢٣ الكويت
Al Kharafi Tower (Building 6) - 1st Floor, Hamad Al Sagr Street, Block 14, Qibla, Kuwait Tel. Off.: 22323856/57 Fax 22323864
C.R.: 27468 - P.O. Box: 26267 Safat, 13123 Kuwait, E-mail: med@warbamed.com, www.warbamed.com

FIELD SERVICE REPORT

Service Report No: 049795	Date: 16/09/24 Quarter: ☐ I ☐ II ☐ III ☐ IV
Customer: AL ORF HOSPITAL	Equipment Location: NICU
Contract No:	
Equipment Name: VENTILATOR	
Manufacturer: HAMILTON	Model: T1
Serial Number: 5627	
Type of Job ☐ Preventive Maintenance ☐ Installation ☐ Warranty Replacement ☒ Emergency Maintenance ☐ Chargeable Repair ☐ Warranty Repair	
Maintenance Carried out and Final Status of the Instrument: checked & found touch screen is not responding. Unit is taken for further evaluation after which the quotation for repair will be given. Also battery needs to be replaced.	

Parts Used

No.:	Part Number:	Description	Quantity

Remarks:

Name: SARAFIAZ Blawday Warba Engineer Signature	Name: Shien End User Signature & Stamp	Authorized Signature & Stamp Biomedical Department
---	--	---