



MEDICAL EQUIPMENT REPAIR REQUEST

No. #9 Date: 16/01/2025

Requested by: Ferre M Williams

Department: NICU Location: 3rd Floor

SR. No.	Machine	Model	Complaint
<u>H0H550195</u>	<u>Medical Incubator</u>	<u>Giraffe</u>	<u>window lock broken</u>

Check within Week ☒ ASAP

Engineer Comments

2 window locks are broken and have
been changed / fixed.

End Comments




Ferre
Requested by

Ahmad Alawadhi
Checked by

Maureen Hassan
Ferre
Approved by

[illegible]

- ③ Service call attended
- ④ Unit checked
- ⑤ Problem: 2 Window lock Broken and have been changed with extra parts they already have.
- ⑥ Unit is in good condition.

REMARKS			
 ATC ENGINEER SIGNATURE  USER SIGNATURE  AUTHORIZED SIGNATURE & STAMP			
JOB COMPLETED	☒ Yes	☐ No	AWAITING PARTS
		☐ Yes	☒ No