

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63517

Customer	Alot Hospital		
Equipment	De fibrillator		
Jahra	Model No.	Reanibex 700	
	Serial No.	200820022414	
Authorized Person	Manufacturer	Breen	
Customer Complaint	Location		
	Contract / Order No.		
Services / Check Carried out	☐ Installation ☐ Warranty		
☒ Inspection / Cleaning	☐ Repair		
☐ Calibration / Performance	☐ On Charge		
☒ Interface / Communication	☐ Maintenance		
☒ Safety Checks	☐ Contract		
☐ Electrical / Electronics	☐ Upgrade		
☒ Accessories Check	☐		
	Time Spent		
Remarks	As per requests, checked the shocks with analyser and it's working. Checked the battery back up, is OK. Performed the handset, foot pedal and paddle interface tests. Checked and found that the unit is working properly.		
No.	Spare Parts Used	Part No.	Qty.
①	50J - Measured value -	52.7 J	
②	100J - Measured value -	101.7 J	
③	200J - Measured value -	203.4 J	
Service Engineer		Authorized Person	
Signature		Signature	
Date		Date	
Customer / User		Signature	
		Date	