

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63852

Customer	Alorj Hospital			Equipment	Patient transfer trolley.		
				Model No.			
				Serial No.	As below		
Authorized Person				Manufacturer			
Customer Complaint	Brake not working			Location	4th floor		
				Contract / Order No.			
Services / Check Carried out	☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check			☐ Installation ☒ Repair ☐ Maintenance ☐ Upgrade	☐ Warranty ☐ On Charge ☐ Contract ☐		
	☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐			Time Spent			
Remarks	Service call attended. we inspect the unit and found that two trolley brake is not working and one trolley head rest is not stable. we replaced the wheels for two trolley and repaired the head rest. Now the three trolleys are working properly.						
No.	Spare Parts Used			Part No.	Qty.		
	Serial no						
	① JN500033720001 wheel replaced.						
	② N/A new trolley. wheel replaced						
	③ N/A head rest repaired						
Service Engineer				Authorized Person		Customer / User	
Signature				Signature		Signature	
Date				Date		Date	