ATCO Technology Co.
Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO . 336218

ATCOTECH

شركة أتكو التكنولوجيه
تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)
فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)
ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت
Email: sales@atco-kuwait.com
www.atco-kuwait.com
س.ت. ٣٣٦٢١٨

SERVICE / ~~INSTALLATION~~ REPORT No.: 63787

Customer Alorf Hospital	Equipment Electro Surgical Unit.
	Model No. SURGILEC MC3
	Serial No. 209
Authorized Person	Manufacturer
Customer Complaint	Location OR
Ground Plate alarm	Contract / Order No.

Services / Check Carried out

- O Inspection / Cleaning
- O Calibration / Performance
- O Interface / Communication
- O Safety Checks
- O Electrical / Electronics
- O Accessories Check
- O Optics Checks
- O Pneumatic Checks
- O Hydraulic Checks
- O Operational Error
- O Mechanical Checks
- O _____

- O Installation
- O Warranty
- ✓ Repair
- O On Charge
- O Maintenance
- O Contract
- O Upgrade
- O _____

Time Spent

Remarks Service call attended.
we inspect the unit and found that the neutral Plate connection inside broken from the connector.
we Solder the cable with connector and fixed it Properly - Checked and confirmed that the unit is working Properly.

No.	Spare Parts Used	Part No.	Qty.
	GLUE applied and cable fixed		
	working now		

4/1/25

Service Engineer	Authorized Person	Customer / User
Nixon ATCO Technology Co.		
Signature	Signature	Signature
Date 29/12/2025	Date	Date