

MEDICAL EQUIPMENT REPAIR REQUEST

No. Emily Thomas Date: 10.06.2025

Requested by _____

Department Radiology Location _____

SR. No.	Machine	Model	Complaint
	Printer Casesteam	6850	Print not coming due to Jam inside Printer
		Day view	

Check within ☐ Week

☐ ASAP

Engineer Comments

PRINTER NOW WORKING .

End Comments

Emily Thomas

Requested by

Checked By

Approved by



FIELD SERVICE REPORT

DEPT: _____

24/ 20644

CUSTOMER

private

EQUIPMENT

vlt vgs dmb

MODEL NO.

Voluson signature 2

DATE

2 / 7 / 2025

JOB NO.

188152

LOCATION

ALORF hospital

SERIAL NO.

VSC 200405

ASSET

CONTRACT

COST CENTER

7000

BILL PARTS / LABOR ☐WARRANTY ☐CONTRACT ☐INSTALLATION ☐OTHERS ☐

ENGINEER

HOURS

Bostan

2

PARTS USED

QTY.	PART NUMBER	SIV	DESCRIPTION
1	55917499		BW Printer option cable set-U
1	LOCUP-D898 X		Sony Digital BW Video printer

PROBLEM, DIAGNOSIS AND ACTION

Parts installed and system up

REMARKS

JOB COMPLETED

☒ Yes☐ No

AWAITING PARTS

☐ Yes☐ No

ATC ENGINEER SIGNATURE

USER SIGNATURE

Alnoor Sarano

AUTHORIZED SIGNATURE & STAMP

مستشفى العرف
ALORF HOSPITAL

www.atc.com.kw