



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 26/03/2025

Requested by _____

Department LDR Location _____

SR. No.	Machine	Model	Complaint
<u>19770208</u>	<u>infusion pump</u>	<u>19770208</u> <u>OPTIMA PT-15</u>	<u>blowing</u>

Check within Week

ASAP

Engineer Comments

End Comments

Paulksham

Requested by

Checked By

Approved by

SERVICE REPORT

1807

CUSTOMER NAME & ADDRESS AL-ORF Hospital		Date: 10/4/2025			
		Equipment: Infusion pump			
		Manufacture: Fresenius			
		Model No: vial optima PT-IS			
		Serial No: 19770208			
FAULT REPORT / SERVICE DETAILS:					
continuous Alarm					
SERVICE CLARIFICATION: ☐ INSTALLATION ☐ WARRANTY ☐ PMS					
☐ OTHERS ☐ ANNUAL MAINTENANCE CONTRACT					
Work Hour	From		To		Total Working Hours
Work Date					Total Working Days
SERVICE PERFORMED					
After the evaluation found the issue with the power supply module. Repaired the power supply module.					
☒ Completed ☐ Under Observation					
CUSTOMER REMARKS:					
Job Satisfactorily Completed		☐ YES ☐ NO		For Cyrix & TSL Company	
Customer Signature:				Engineer Name: Jayahar A	
				Signature: [Signature]	