

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 1970

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60895

Customer	Aleef Hospital Jahra	Equipment	Portable X Ray
		Model No.	Cybermobil TS
		Serial No.	T7296
Authorized Person		Manufacturer	Pomax
Customer Complaint		Location	NECH
No exposure		Contract / Order No.	

Services / Check Carried out

- | | |
|--|--|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ |

- | | |
|--|---|
| ☐ Installation | ☐ Warranty |
| ☒ Repair | ☒ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ |

Time Spent

Remarks

Service call addressed. Checked end board cable from inside the machine for hand switch damaged condition. Repaired the damaged end found it working properly. During exposure sometimes other errors are appearing, this is due to related to

No.	Spare Parts Used	Part No.	Qty.
	main CPU board.		

Service Engineer

ATCO Technology Co.

Signature

Date

28/09/2024

Authorized Person

Signature

Date

Aji

Customer / User

Signature

Date

Aji
29/09