

1473

Roche for the Trade in Medicines, Equipment, Devices
and Medical Supplies SPC
Customer Report Form



Case number CAS-0019956721

Order number WO-27726332 | Visit Date 03/25/2024

Instrument Details

Instrument/Module cobas b 221<6>=Roche OMNI
S6 system

Serial Number 19239

Additional Details

PO

Customer Details

Customer Number 0050250925

Customer Name Dr. Mansour Alorf Hospital

Street Address Mansour Saad Al Bannaq
Street

Zip 01025 - Jahra

Contact Name Gina Bisenio

Contact Phone +96565991613

Service Activity Code Service Visit

Purpose of Visit

CUSTOMER ALLEGATION

=====

Created by: Gina Bisenio, biomed@alorf.com (+96565991613)

Preferred callback time:

Instrument: b 221 6 Roche OMNI S6 system (Blood Gas/Electrolytes) (Serial: 19239)

Alarm code: not-applicable

Issue type: OTHER

Lot no/Vendor: not applicable

Initial description: by pass wash error 2

Issue timeline:

1: Issue created (24.03.2024 16:18 GMT+00:00)

ACTIONS

=====

Type your descriptive analysis here

CONCLUSION

=====

Log a call or note

Performed Activities

Guided customer by audio and video calls to clean the t&d disk and perform wash Cycles.

Machine working.

Time Report

Category	Start	End	Hours
05913616001-Service labour time	03/25/2024 10:00 am	03/25/2024 11:00 am	1.00

Customer
Gina Bisenio

Roche Support
Sunash Shukoor

A hand-drawn sketch of a helical spring, showing a coiled wire with several loops. The drawing is done in black ink on a white background.

- Roche Support person confirms the above mentioned work and checks have been completed..
- Customer representative confirms they have reviewed the work carried out as described and accepts the system for use.
- Thank you for your trust.

Address: Al Mirqab, Al Soor Street, Block 1, Burj Al Shaya, Plot 6, Floor 5, Office
11. P.O. Box: 467, Safat - 13005 Kuwait
Hotline: 25770360
Fax: 22059262

Customer Assistance: