

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63764

Customer	AL-Oz Hospital Jahra.	Equipment	Lab Water Bath.
		Model No.	1002.
		Serial No.	11304615J
Authorized Person		Manufacturer	GFL.
Customer Complaint	Not Working.	Location	Lab.
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	Checked and found thermostat tripped and reset the thermostat. Found the unit is working properly.		
* Please make sure that minimum water level has to check.			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
CHERNAN PHILIP.		Hany W2 Hanna	
ATCO Technology Co.		د. هاني وليم زكي حنا	
Signature	Signature	طبيب مسجل أول بيوكمستري	
Date	Date	مستشفى العرف	
		22025120371	