

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 1775

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجاهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61019

Customer	Alex Hospital Jahra	Equipment	Mamogram Unit								
		Model No.	Melioranthuc								
		Serial No.	HIELC/0063/C1								
Authorized Person		Manufacturer	Metalbonca								
Customer Complaint		Location	X-Ray								
		Contract / Order No.									
Services / Check Carried out	<table><tr><td>☐ Installation</td><td>☐ Warranty</td></tr><tr><td>☒ Repair</td><td>☐ On Charge</td></tr><tr><td>☐ Maintenance</td><td>☐ Contract</td></tr><tr><td>☐ Upgrade</td><td>☐ _____</td></tr></table>			☐ Installation	☐ Warranty	☒ Repair	☐ On Charge	☐ Maintenance	☐ Contract	☐ Upgrade	☐ _____
☐ Installation	☐ Warranty										
☒ Repair	☐ On Charge										
☐ Maintenance	☐ Contract										
☐ Upgrade	☐ _____										
☐ Inspection / Cleaning	☐ Optics Checks										
☐ Calibration / Performance	☐ Pneumatic Checks										
☐ Interface / Communication	☐ Hydraulic Checks										
☐ Safety Checks	☐ Operational Error										
☐ Electrical / Electronics	☐ Mechanical Checks										
☐ Accessories Check	☐ _____										
		Time Spent									

Remarks

After ~~overhaul~~ overlooked by manufacturer. Recommended to change the spare parts of compression board and controller unit. will be giving quotation for replacements

No.	Spare Parts Used	Part No.	Qty.

Service Engineer ATCO Technology Co. Signature Date 16/10/2024	Authorized Person Signature Date	Customer / User Signature Date
---	--	--