

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



logno-2588

شركة أتكوتكنولوجيه

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61992

Customer	Alex Hospital.			Equipment	CSSD washer.		
				Model No.	Uniclean PL II		
				Serial No.	230506		
Authorized Person				Manufacturer	mmm		
Customer Complaint				Location	CSSD		
unloading door not opening error: Time exceeding				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation		☐ Warranty	
				☒ Repair		☐ On Charge	
				☐ Maintenance		☐ Contract	
				☐ Upgrade		☐	
				Time Spent			
Remarks Service call attended. we inspect the unit and found that the door is not opening. we manually open the door and reset the error. we checked the relay and its working. we tested one cycle and now the unit is working properly.							
No.	Spare Parts Used			Part No.		Qty.	
	under warranty						
Service Engineer Nixon / nasa ATCO Technology Co. Signature Date 29/08/2024		Authorized Person 29/8/24 Signature Date		Customer / User Signature Date			