



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 01-10-2024

Requested by Jharna Rai

Department Pediatrics Location 4th Floor

SR. No.	Machine	Model	Complaint
01-05-0144	vital sign monitor	Cas740	Not charging.

Check within

☐ Week

☐ ASAP

Engineer Comments

End Comments

Jharna Rai

Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63156

Customer	Alorj Hospital	Equipment	Patient Monitor
		Model No.	CAS 740
		Serial No.	01050144
Authorized Person		Manufacturer	Caemed.
Customer Complaint		Location	PEDIA

Not working	Contract / Order No.
	☐ Installation ☐ Warranty ☐ Repair ☐ On Charge ☐ Maintenance ☐ Contract ☐ Upgrade ☐
Services / Check Carried out	Time Spent
☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐	

Remarks	Battery defect in not charging. will check availability of Battery. Unit delivered to Can be used without battery
---------	---

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Signature Date	Authorized Person Signature Date	Customer / User Signature Date
 ATCO Technology Co. 3/10/24	 8/10/24	