

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2161

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61993

Customer	Aloof Hospital			Equipment	Remote for OT Bed.		
				Model No.	Nuvo 750		
				Serial No.			
Authorized Person				Manufacturer	Nuvo.		
Customer Complaint	cable broken.			Location	OR		
				Contract / Order No.			
Services / Check Carried out	☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check			☐ Installation ☐ Repair ☐ Maintenance ☐ Upgrade	☐ Warranty ☐ On Charge ☐ Contract ☐		
	☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐			Time Spent			
Remarks	we inspect the remote and found that the cable is broken. we taken the remote for inspection in our workshop. once we check the remote and update if it is repairable or not.						
No.	Spare Parts Used			Part No.	Qty.		
	taken for repair						
Service Engineer	Authorized Person			Customer / User			
Nixon/masa ATCO Technology Co.							
Signature	Signature			Signature			
Date 28/08/2024	Date			Date			