

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61086

Customer	Al orf Hospital			Equipment	Beef Patient Trolley		
				Model No.			
				Serial No.	DN500033780001		
Authorized Person				Manufacturer			
Customer Complaint				Location	4th Floor		
The trolley doesn't stop (stop leg)				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
				Time Spent			
Remarks A service call attended we inspect the trolley, we found that there is one iron volve not in its place. we adjust the iron volve. and we checked all the screws and find and found there is one screw not missing. we checked the trolley and it's working properly, one side.							
No.	Spare Parts Used			Part No.	Qty.		
	"not working"						
Service Engineer Nixon / Mousa ATCO Technology Co. Signature Date 4/9/2024		Authorized Person Signature Date		Customer / User Signature Date			