

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



Log no - 1746

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب. ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60609

Customer	Al- Ory Hospital Kuwait	Equipment	Blood Bank Refrigerator																								
		Model No.																									
		Serial No.	FFCG20DS K022 L1																								
Authorized Person		Manufacturer	Desmem																								
Customer Complaint		Location	LAB																								
	No Cooling	Contract / Order No.																									
Services / Check Carried out	<table><tr><td>☐ Inspection / Cleaning</td><td>☐ Optics Checks</td><td>☐ Installation</td><td>☐ Warranty</td></tr><tr><td>☐ Calibration / Performance</td><td>☐ Pneumatic Checks</td><td>☐ Repair</td><td>☐ On Charge</td></tr><tr><td>☐ Interface / Communication</td><td>☐ Hydraulic Checks</td><td>☐ Maintenance</td><td>☐ Contract</td></tr><tr><td>☐ Safety Checks</td><td>☐ Operational Error</td><td>☐ Upgrade</td><td>☐ _____</td></tr><tr><td>☐ Electrical / Electronics</td><td>☐ Mechanical Checks</td><td></td><td></td></tr><tr><td>☐ Accessories Check</td><td>☐ _____</td><td></td><td></td></tr></table>			☐ Inspection / Cleaning	☐ Optics Checks	☐ Installation	☐ Warranty	☐ Calibration / Performance	☐ Pneumatic Checks	☐ Repair	☐ On Charge	☐ Interface / Communication	☐ Hydraulic Checks	☐ Maintenance	☐ Contract	☐ Safety Checks	☐ Operational Error	☐ Upgrade	☐ _____	☐ Electrical / Electronics	☐ Mechanical Checks			☐ Accessories Check	☐ _____		
☐ Inspection / Cleaning	☐ Optics Checks	☐ Installation	☐ Warranty																								
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Repair	☐ On Charge																								
☐ Interface / Communication	☐ Hydraulic Checks	☐ Maintenance	☐ Contract																								
☐ Safety Checks	☐ Operational Error	☐ Upgrade	☐ _____																								
☐ Electrical / Electronics	☐ Mechanical Checks																										
☐ Accessories Check	☐ _____																										
Remarks	Checked system, note low coolant gas and checked filter + Fan Refilled R404 gas + Replaced Filter & Compressor fan. Checked operations found OK																										
No.	Spare Parts Used	Part No.	Qty.																								
Service Engineer	Authorized Person	Customer / User																									
Alio team		Mr. 30/8/24																									
ATCO Technology Co.																											
Signature	Signature	Signature																									
Date	Date	Date																									
25-5-24																											