

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61089

Customer	Al orf Hospital			Equipment	Auto Clave		
				Model No.	sterilClave		
				Serial No.	18x25 721		
Authorized Person				Manufacturer	COMINOX		
Customer Complaint				Location	6th floor (dentist)		
Bowditch Test, colour not changing				Contract / O	No.		
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
				Time Spent			
Remarks Service call attended we inspect the unit and tried vacuum test and the pressure variation was 3mbar, Leakage is less compared with actual one, 0.3 mbar/min. we tried the Bowditch test with Test Kit, and the colour is changing we keeping the unit under observation for 2 days we run a Bowditch cycle and it's working properly.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer		Authorized Person		Customer / User			
Moussa				د. عبد الرحمن السعيد عبد الله			
ATCO Technology Co.				طبيب أسنان مسجل أول الاستشارة الصناعية			
Signature		Signature		Signature			
Date 9/9/2024		Date		Date			