

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60672

Customer	Aloof Hospital	Equipment	Dental Autoclave
		Model No.	sterilizer 24 Slot
		Serial No.	18 x 25721
Authorized Person		Manufacturer	Cominox
Customer Complaint		Location	Dental Lab
	No problem	Contract / Order No.	

Services / Check Carried out

- | | |
|--|--|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ |

- | | |
|---------------------------------------|------------------------------------|
| ☐ Installation | ☐ Warranty |
| ☐ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ |

Time Spent

Remarks: Service call attended - Found fuse blown inside. Replaced fuse, tested cycle, unit is working ok

No.	Spare Parts Used	Part No.	Qty.
	Fuse 16A		1

Service Engineer

ATCO Technology Co.

Signature

Date

Authorized Person

Signature

Date

Customer / User

Signature

Date

د. هادي الرحمن السعيد عبدربه
طبيب اسنان مسجل اول الاستعاضة الصناعية
مستشفى السيف
22019100056