



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 23/02/25

Requested by Reema

Department LDR Location _____

SR. No.	Machine	Model	Complaint
19357626	infusion pump		not working

Check within ☐ Week

☐ ASAP

Engineer Comments

End Comments


Reema
Requested by

Checked By

Approved by

SERVICE REPORT

1780

CUSTOMER NAME & ADDRESS Al ORF Hospital		Date:		24/2/2025	
		Equipment:		Infusion pump	
		Manufacture:		Fresenius	
		Model No:		optima PT IS	
		Serial No:		19357626	
FAULT REPORT / SERVICE DETAILS:					
Error 43					
SERVICE CLARIFICATION: ☐ INSTALLATION ☐ WARRANTY ☐ PMS					
☒ OTHERS ☐ ANNUAL MAINTENANCE CONTRACT					
Work Hour	From		To		Total Working Hours
Work Date					Total Working Days
SERVICE PERFORMED					
After the evaluation found the unit showing continuous alarm error 43. need to take the unit to our company for repair.					
☐ Completed ☒ Under Observation					
CUSTOMER REMARKS:					
Job Satisfactorily Completed ☐ YES ☐ NO			For Cyrix & TSL Company		
Customer Signature: 			Engineer Name: Jawahar A		
			Signature: 		