

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)
فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)
ص.ب : ٢٣٥٥ الجهراء , ٠١٠٢٥ - الكويت
Email: sales@atco-kuwait.com
www.atco-kuwait.com
س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63830

Customer	ALORF Hospital		Equipment	ECG	
			Model No.	CARDIPIA800H	
			Serial No.	C800152102	
Authorized Person			Manufacturer	TRISMED	
Customer Complaint			Location	1M 1st Floor	
V2 V3 leads not detecting			Contract / Order No.		
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐			☐ Installation		☐ Warranty
			☐ Repair		☐ On Charge
			☐ Maintenance		☐ Contract
			☐ Upgrade		☐ <u>EM</u>
			Time Spent		
Remarks					
Service call attended, we inspected the unit and tested the ECG cable with stimulator. Found lead ends are dirty. We cleaned the leads and replaced the lead connectors from the store. All connectors are tested ok. System working good.					
No.	Spare Parts Used		Part No.	Qty.	
Service Engineer		Authorized Person		Customer / User	
Nixon & Pine				Hussein Hassan	
ATCO Technology Co.				Signature	
Signature		Signature		Date	
Date 06/02/2025		Date		9-2-25	