

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 1735

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60502

Customer	Alert Hospital	Equipment	Centrifuge
		Model No.	Z 206A
		Serial No.	601900023
Authorized Person		Manufacturer	Hermle
Customer Complaint		Location	laboratory.
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
		☒ Repair	☐ On Charge
☐ Inspection / Cleaning		☐ Optics Checks	
☐ Calibration / Performance		☐ Pneumatic Checks	
☐ Interface / Communication		☐ Hydraulic Checks	
☐ Safety Checks		☐ Operational Error	
☐ Electrical / Electronics		☐ Mechanical Checks	
☐ Accessories Check		☐ _____	
		Time Spent	
Remarks			
We inspect the unit and found that the power switch is faulty. We replaced the power switch with new one. Now the unit is working. checked and confirmed that the unit is working properly.			
No.	Spare Parts Used	Part No.	Qty.
1.	power switch Hermle.		1
Service Engineer		Authorized Person	
Nixon		Farida	
ATCO Technology Co.			
Signature		Signature	
Date 20/06/2024.		Date 20/6/24	
Customer / User			