

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58937

Customer	Aloof Hospital			Equipment	ECG		
				Model No.	SE 1200 Express		
				Serial No.	92300 006009		
Authorized Person				Manufacturer	Eden		
Customer Complaint				Location	ER		
	NO trace on G			Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
				Time Spent			
Remarks							
Service Call attended.							
we inspect the ECG unit and checked the ECG using stimulator. we found that the V2 port is not trace. The cable is faulty.							
Replaced the ECG cable and checked and confirmed that the unit is working properly							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer		Authorized Person		Customer / User			
Mison							
ATCO Technology Co.							
Signature		Signature		Signature			
Date		Date		Date			
17/03/2024		19/3/2024					