

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب. ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63672

Customer	ALORF Hospital Jahra	Equipment	ENT Work station		
		Model No.	Futuraent 3		
		Serial No.	111632045		
Authorized Person		Manufacturer	Entimed		
Customer Complaint		Location	ENT		
	Water leaking	Contract / Order No.			
Services / Check Carried out		☐ Installation	☐ Warranty		
		☐ Repair	☐ On Charge		
		☐ Maintenance	☐ Contract		
		☐ Upgrade	☐ _____		
		Time Spent			
Remarks					
Cleaned & Thermo full cryo. water tubes					
Cleaned drain systems					
Tested and confirmed all unit working properly					
System working good					
No.	Spare Parts Used	Part No.	Qty.		
Service Engineer		Authorized Person			
JP & Jirree					
ATCO Technology Co.		مستشفى الجهر 22023120704			
Signature		Signature			
Date		Date			
22/05/20					
Customer / User		Signature			
		Date			