

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكوتكنولوجيه

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63668

Customer	Alort Hospital			Equipment	Patient monitor module		
				Model No.	Xm1 module		
				Serial No.	M19711220001		
Authorized Person				Manufacturer	Eden		
Customer Complaint				Location	OR		
				Contract / Order No.			
Services / Check Carried out				☐ Installation ☐ Warranty			
				☒ Repair ☐ On Charge			
☐ Inspection / Cleaning ☐ Optics Checks				☐ Maintenance ☐ Contract			
☐ Calibration / Performance ☐ Pneumatic Checks				☐ Upgrade ☐ _____			
☐ Interface / Communication ☐ Hydraulic Checks							
☐ Safety Checks ☐ Operational Error							
☐ Electrical / Electronics ☐ Mechanical Checks							
☐ Accessories Check ☐ _____							
				Time Spent			
Remarks							
we replaced the main board with new one. checked and confirmed that the Xm1 module is working properly.							
No.	Spare Parts Used			Part No.		Qty.	
	working						
	20/5/25						
Service Engineer				Authorized Person		Customer / User	
Nixon / JP							
ATCO Technology Co.							
Signature				Signature		Signature	
Date 18/05/2025				Date		Date	