

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63519

Customer	Alseef Hospital Jahra		Equipment	Defibrillator	
			Model No.	Reanibex 700	
			Serial No.	200820022213	
Authorized Person			Manufacturer	Bexen	
Customer Complaint			Location	NICU	
			Contract / Order No.		
Services / Check Carried out ☒ Inspection / Cleaning ☐ Calibration / Performance ☒ Interface / Communication ☒ Safety Checks ☐ Electrical / Electronics ☒ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☒ Operational Error ☐ Mechanical Checks ☐			☐ Installation	☐ Warranty	
			☐ Repair	☐ On Charge	
			☐ Maintenance	☐ Contract	
			☐ Upgrade	☐	
			Time Spent		
Remarks	As per request, checked the shock delivered with analyzer. and it working. Performed Hasekware, front panel and paddle interface test. Checked and found that the unit is working properly				
No.	Spare Parts Used		Part No.	Qty.	
①	50J - Measured value - 46.3J				
②	100J - Measured value - 99.9J				
③	200J - Measured value - 194J				
Service Engineer		Authorized Person		Customer / User	
Smith ATCO Technology Co.		Fene Joy Williams			
Signature		Signature		Signature	
Date 27/11/2024		Date		Date	