



MEDICAL EQUIPMENT REPAIR REQUEST

No. # 11 Date: 13/01/25

Requested by: Ferre Jny Williams

Department: NICU Location: 3rd Floor

SR. No.	Machine	Model	Complaint
<u>HDT50691</u>	<u>Medical Incubator</u>	<u>Giraffe</u>	<u>Pedal pad broken</u>

Check within Week ☒ ASAP ☐

Engineer Comments

The Pedal Pad is Broken and need to Change
it with new one. The Unit is functioning
But cover Broken.

End Comments

Ferre
Requested by

Ahmad Alawadhi
Checked by



Waleed
Approved by

FIELD SERVICE REPORT

LCS 12903

[illegible]

- ⊕ service call attended
- ⊕ unit checked
- ⊕ Problem: Pedal Pad is functioning well and its working. But, The cover of the Pad is Broken and need to Bring a new Pad
- ⊕ Repair Request is needed

REMARKS	
JOB COMPLETED ☐ Yes ☒ No	AWAITING PARTS ☒ Yes ☐ No
ATC ENGINEER SIGNATURE 	USER SIGNATURE  مستشفى القوف ALORF HOSPITAL
www.atc.com.kw	