

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60851

Customer	Alfah Hospital Jahra		Equipment	Mammogram Unit	
			Model No.	Heliocision C	
			Serial No.	HILLC/0063/C1	
Authorized Person			Manufacturer	Metalbionica	
Customer Complaint			Location	X-Ray	
System Error X-Ray sequence aborted Apply compression first			Contract / Order No.		
Services / Check Carried out			☐ Installation	☐ Warranty	
☐ Inspection / Cleaning	☐ Optics Checks		☐ Repair	☐ On Charge	
☐ Calibration / Performance	☐ Pneumatic Checks		☐ Maintenance	☐ Contract	
☐ Interface / Communication	☐ Hydraulic Checks		☐ Upgrade	☐ _____	
☐ Safety Checks	☐ Operational Error		Time Spent		
☐ Electrical / Electronics	☐ Mechanical Checks				
☐ Accessories Check	☐ _____				
Remarks	Service call attended. Checked and found the compressor is not detecting even is compressed condition. Need technical support from the manufacturer				
No.	Spare Parts Used	Part No.	Qty.		
Service Engineer	Authorized Person	Customer / User			
Signature Date 29/9/2024	Signature Date	Signature Date			