

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.		
PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____			
Date of Occurrence: <u>07.05.2025</u>		Time of Occurrence: <u>0500H</u>	Occurrence Location: <u>LDR</u>
If event involved employee - Staff Name: <u>Neetha Rabul</u>		Age: <u>35</u> Sex: <u>fe</u> Emp#: <u>3826</u>	
If event involved patient - Patient Name: _____		Age: _____ Sex: _____ File#: _____	
If event involved Visitor - Visitor Name: _____		Phone No.: _____	
Address: _____			

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR ☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	SECURITY VARIANCE ☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	EQUIPMENT/SUPPLIES ☐ Electrical Problem ☒ Not Working ☐ Improper Use ☐ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>Infusion pump</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)	Response to Treatment/Outcome:	
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: <u>Neetha Rabul</u>	ID No.	Dept. <u>LDR</u>	Position
IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication				

Assessment/ Diagnosis		
Physician Name:	Dept.:	ID No.:

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

07.05.2025 0500H.

Infusion pump receive from Bio-med, this machine connected with patient, @ 0500H. that time it was showing error in display. I removed from patient and not working tag attached.

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ **Moderate.** Increased length of Stay. Increased level of care.
- ☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT	REFERRED TO QUALITY MANAGEMENT DEPT.
✓ channeling was done ✓ advised the steps to use the machine & check if its properly working before receiving for the Biomed. Testing it later if to be used by patient and finding out not working is not acceptable. Will cause delay in treating our patient → referred to Biomed the machine	- Infusion pump error noted & device replaced promptly with no treatment delay. Unit tagged & referred to Biomedical for assessment. - Awaiting report on findings & corrective action.
Dept. Head: Christina Owen Date: 8/5/25	Quality Mgmt. Dept. Date: Charlotte T.B / pd
ID No. 1022 Department: Humsu	Comments:

Department Managers: Forward Completed forms to Quality Management Department