

ATCO Technology Co.

Tel: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة اتكو التكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 59177

Customer	Alor Hospital	Equipment	Pympnanonm
		Model No.	Viola Plm.
		Serial No.	1M 1DD17203029
Authorized Person		Manufacturer	Arbomd
Customer Complaint		Location	ENR.
	Calibration clean	Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ <u>Em</u>
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____	Time Spent	
Remarks	re-calibrated pressure transducer and unit. tested unit is working ok.		

No.	Spare Parts Used	Part No.	Qty.

Service Engineer ATCO Technology Co. Signature Date 12/3/24	Authorized Person 12/3/24 Signature Date	Customer / User Signature Date
--	---	--