

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية
1552-24567725

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58341

Customer	Alor Hospital														
Equipment	ECG														
Model No.	SE 1200 Express.														
Serial No.	460016-M21C11560001														
Authorized Person	Edam														
Customer Complaint	Printer not working.														
Services / Check Carried out	<table><tr><td>☐ Inspection / Cleaning</td><td>☐ Optics Checks</td></tr><tr><td>☐ Calibration / Performance</td><td>☐ Pneumatic Checks</td></tr><tr><td>☐ Interface / Communication</td><td>☐ Hydraulic Checks</td></tr><tr><td>☐ Safety Checks</td><td>☐ Operational Error</td></tr><tr><td>☐ Electrical / Electronics</td><td>☐ Mechanical Checks</td></tr><tr><td>☐ Accessories Check</td><td>☐</td></tr></table>			☐ Inspection / Cleaning	☐ Optics Checks	☐ Calibration / Performance	☐ Pneumatic Checks	☐ Interface / Communication	☐ Hydraulic Checks	☐ Safety Checks	☐ Operational Error	☐ Electrical / Electronics	☐ Mechanical Checks	☐ Accessories Check	☐
☐ Inspection / Cleaning	☐ Optics Checks														
☐ Calibration / Performance	☐ Pneumatic Checks														
☐ Interface / Communication	☐ Hydraulic Checks														
☐ Safety Checks	☐ Operational Error														
☐ Electrical / Electronics	☐ Mechanical Checks														
☐ Accessories Check	☐														
Contract / Order No.	<table><tr><td>☐ Installation</td><td>☐ Warranty</td></tr><tr><td>☒ Repair</td><td>☐ On Charge</td></tr><tr><td>☐ Maintenance</td><td>☐ Contract</td></tr><tr><td>☐ Upgrade</td><td>☐</td></tr></table>			☐ Installation	☐ Warranty	☒ Repair	☐ On Charge	☐ Maintenance	☐ Contract	☐ Upgrade	☐				
☐ Installation	☐ Warranty														
☒ Repair	☐ On Charge														
☐ Maintenance	☐ Contract														
☐ Upgrade	☐														
Time Spent															
Remarks	Service call attended. when we inspect the unit we found that the printer cover is broken. Need to replace the printer cover.														
No.	Spare Parts Used	Part No.	Qty.												
Service Engineer	Authorized Person	Customer / User													
Miscon															
ATCO Technology Co.															
Signature	Signature	Signature													
Date	Date	Date													
29/02/2024	4/3/2024														