

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.		
PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____			
Date of Occurrence: <u>9/6/25</u>		Time of Occurrence: <u>8:45 AM</u>	
		Occurrence Location: <u>CSSD</u>	
If event involved employee - Staff Name: <u>Ramge</u>		Age: _____	Sex: _____ Emp#: <u>4231</u>
If event involved patient - Patient Name: _____		Age: _____	Sex: _____ File#: _____
If event involved Visitor - Visitor Name: _____		Phone No.: <u>232</u>	
Address: _____			

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☒ Not Working ☐ Improper Use ☐ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☒ Other <u>CSSD WASHER</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE		
☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: _____	ID No. _____	Dept. _____	Position _____
IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No. Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication				
Assessment/ Diagnosis				
Physician Name:	Dept.:		ID No.:	

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

Respected Madam Today Morning. I start- washer II
Disinfectant washing pot cycle finish. But- does not open.

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ **Moderate.** Increased length of Stay. Increased level of care.
- ☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
7 mth to PIONEER		- Referred to Biomedical for repair.	
Dept. Head: <i>Christina Carter</i>	Date: <i>9/6/2020</i>	Quality Mgmt. Dept. <i>Charlotte TB (RN)</i>	Date:
ID No. <i>1022</i>	Department: <i>NURSING</i>	Comments:	

Department Managers: Forward Completed forms to Quality Management Department



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 9/6/25

Requested by Remye

Department CSSD Location CSSD

SR. No.	Machine	Model	Complaint
<u>230506</u>	<u>Washer no 2.</u>	<u>uniclean</u>	<u>does not open.</u>

Check within ☐ Week

☐ ASAP

Engineer Comments

End Comments

Remye
Requested by

Checked By

Approved by

Customer <u>Atcof Ho Spital</u>		Equipment <u>Washer/Difingeln</u>	
		Model No. <u>Unitclean</u>	
		Serial No. <u>230506</u> (#2)	
Authorized Person		Manufacturer <u>mmm</u>	
Customer Complaint		Location <u>CSSD</u>	
<u>unloading door not opened</u>		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐ _____		☐ Installation ☐ Warranty	
		☐ Repair ☐ On Charge ☐ Maintenance ☐ Contract ☐ Upgrade ☐ <u>EM</u>	
Remarks <u>Inspected unit, manual reset done for unloading side door relays and Pies. Run full cycle, confirmed unit working properly.</u>		Time Spent	
No.	Spare Parts Used	Part No.	Qty.
	<u>opened now</u>		
	<u>working</u>		
	<u>11/6/25</u>		
Service Engineer <u>ATCO Technology Co.</u>		Authorized Person	
Signature <u>ATCO</u>		Signature	
Date <u>9/6/25</u>		Date	
Customer / User		Signature	
		Date	