

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو للتكنولوجيا

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 59776

Customer	Alort Hospital			Equipment	Centrifuge		
				Model No.	Z 206A		
				Serial No.	60130506		
Authorized Person				Manufacturer	HERMLE		
Customer Complaint	High sound while using.			Location	Laboratory		
				Contract / Order No.			
Services / Check Carried out				☐ Installation		☐ Warranty	
				☒ Repair		☐ On Charge	
☐ Inspection / Cleaning				☐ Optics Checks		☐ Contract	
				☐ Calibration / Performance		☐ Pneumatic Checks	
☐ Interface / Communication				☐ Hydraulic Checks		☐ Upgrade	
				☐ Safety Checks		☐ Operational Error	
☐ Electrical / Electronics				☐ Mechanical Checks		☐	
				☐ Accessories Check		☐	
Remarks				Time Spent			
Service call attended							
we inspect the unit and found that some paper							
pieces are inside the rotor part.							
we removed the papers and checked the unit.							
checked and confirmed that the unit is working							
properly.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer		Authorized Person		Customer / User			
Nixon				Dr. S. 24/4			
ATCO Technology Co.				2024			
Signature		Signature		Signature			
Date		Date		Date			
18/04/2024							