

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



log no - 1775

شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 54422

Customer	Alor Hospital Jahra			Equipment	Mamogram Unit				
				Model No.	Helianthus C.				
				Serial No.	HSLC/0063/0C1				
Authorized Person				Manufacturer	Metalbonica				
Customer Complaint				Location	X Ray Dept				
				Contract / Order No.					
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐ _____				☐ Installation ☒ Repair ☐ Maintenance ☐ Upgrade			☐ Warranty ☐ On Charge ☐ Contract ☐ _____		
				Time Spent					
Remarks									
As per order, Replaced the following parts with new one. Performed the daily QC test and Passed. Done phantom tests. Checked all the program settings. Checked and found that the unit is working in good condition.									
No.	Spare Parts Used			Part No.	Qty.				
①	Compression board PCB16-269				1				
②	Control board PCB14-240				1				
Service Engineer			Authorized Person		Customer / User				
Signature: <i>Sujith + Jaya</i> ATCO Technology Co.									
Signature: <i>[Signature]</i>			Signature: <i>[Signature]</i>		Signature: <i>[Signature]</i>				
Date: 07/11/2024			Date:		Date: Nov 7, 2024				