

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 1770

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 59796

Customer	Al ost 170 spiral	Equipment	Center fuge
		Model No.	Z 206 A
		Serial No.	60900024
Authorized Person		Manufacturer	Heemle
Customer Complaint		Location	
		Contract / Order No.	

Services / Check Carried out

- | | |
|---|---|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ _____ |

- | | |
|------------------------------------|---------------------------------|
| ☐ Installation | ☐ Warranty |
| ☐ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ _____ |

Time Spent

Remarks

Lid lock and main power transformer faulty. / and need replacement. We will give quotation for repairs at the earliest

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

[Signature]
ATCO Technology Co.

Signature

Date

9/10/24

Authorized Person

Signature

Date

Customer / User

[Signature]
Dr. Aayat Samy

Signature

Date

10-10-2024

د. ايأت سامي محمد عبدالله
طبيب مسجل باثولوجي
مستشفى العرف
22023120560