

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجي

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63002

Customer	Alesf Hospital Hospital			Equipment	Blood Analyzer Ref:
	Jahra			Model No.	DS-FS30E
				Serial No.	FF0920PSK0221
Authorized Person				Manufacturer	Desmar Scientific
Customer Complaint				Location	Lab
				Contract / Order No.	
Services / Check Carried out				☐ Installation ☐ Warranty	
				☒ Repair ☒ On Charge	
☐ Inspection / Cleaning ☐ Optics Checks				☐ Maintenance ☐ Contract	
☐ Calibration / Performance ☐ Pneumatic Checks				☐ Upgrade ☐	
☐ Interface / Communication ☐ Hydraulic Checks				Time Spent	
☐ Safety Checks ☐ Operational Error					
☐ Electrical / Electronics ☐ Mechanical Checks					
☐ Accessories Check ☐					
Remarks					
Replacing the circulator pen with new checked and c found that the unit is working properly.					
No.	Spare Parts Used			Part No.	Qty.
Service Engineer		Authorized Person		Customer / User	
Signature		Signature		Dr. Aafat Samy	
ATCO Technology Co.				29-9-2024	
Signature		Signature		Signature	
Date		Date		Date	

د. آيات سامي محمد عبد الله
طبيب مسجل باثولوجي
مستشفى الصرف
22023120560