

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO . 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)
فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)
ص.ب: ٢٣٥٥ الجهره، ٠١٠٢٥ - الكويت
Email: sales@atco-kuwait.com
www.atco-kuwait.com
س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60724

Customer	Aloft Hospital			Equipment	Dermas laser machine		
				Model No.	SP Dynamic		
				Serial No.	16100425		
Authorized Person				Manufacturer	Fotona		
Customer Complaint	Water leakage			Location	4th floor DERMA		
				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
				Time Spent			
Remarks we replaced the pump connector. with new one. connected the tubing as its normal way. Tested the leakage. Refill the distilled water. al found that there is no leakage. we checked the power and everything is in the limit. checked and confirmed that the unit is working properly.							
No.	Spare Parts Used			Part No.	Qty.		
1	Pump metal connector			FOTO304-97563	2		
2	Distilled water				3 btl		
flash lamp count rd 4ag:- 20574 Er 4ag :- 113186							
Service Engineer Niscom ATCO Technology Co. Signature Date 06/08/2024			Authorized Person Signature Date			Customer / User طبيب بشري مسجل جلالية وناسلية 78215877 مستشفى الشرف Signature Date	