

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.		
PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____			
Date of Occurrence: <u>14/06/25</u>		Time of Occurrence: <u>9 am</u>	
Occurrence Location: <u>PT female.</u>			
If event involved employee - Staff Name: <u>Praveena T.</u>		Age: <u>30</u> Sex: <u>F</u> Emp#: <u>4270</u>	
If event involved patient - Patient Name: _____		Age: _____ Sex: _____ File#: _____	
If event involved Visitor - Visitor Name: _____		Phone No.: _____	
Address: _____			

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☒ Not Working ☐ Improper Use ☐ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>Pedal of bed is not working</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)	Response to Treatment/Outcome:	
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: <u>Praveena T</u>	ID No. <u>4270</u>	Dept. <u>PT</u>	Position <u>Physiotherapist</u>
IMMEDIATE ACTION				
Physician Notified ☐ Yes ☐ No. Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication				

Assessment/ Diagnosis			
Physician Name:	Dept.:	ID No.:	

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

pedel of one bed is not working properly.
Already this bed is repaired last month.

PART D. Severity Codes (Please check one).


- ☐ Minor. No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ Moderate. Increased length of Stay. Increased level of care.
- ☐ Major. Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ Sentinel/Catastrophic. Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ Near Miss. For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
referred to Biomed		- Recommend reassessment or replacement. Referred to Biomedical team.	
recently repaired, but now pedal not working			
to call the company			
Dept. Head: CHRISTINA CUEVA		Quality Mgmt. Dept. Charlotte TB / RN	
Date: 14/6/2015		Date:	
ID No. 1022		Comments:	
Department: Nursing			

Department Managers: Forward Completed forms to Quality Management Department

Customer ALORF HOSPITAL		Equipment Medicated Electric bed	
Jahra		Model No.	
		Serial No. EO130018	
Authorized Person		Manufacturer Medisa	
Customer Complaint		Location Physiotherapy	
Bed control Switch not working		Contract / Order No.	
Services / Check Carried out ☒ Inspection / Cleaning ☐ Optics Checks ☒ Calibration / Performance ☐ Pneumatic Checks ☒ Interface / Communication ☐ Hydraulic Checks ☒ Safety Checks ☐ Operational Error ☒ Electrical / Electronics ☐ Mechanical Checks ☒ Accessories Check ☐ _____		☐ Installation ☐ Warranty	
		☐ Repair ☐ On Charge ☐ Maintenance ☐ Contract ☐ Upgrade ☒ EM	
		Time Spent	
Remarks Checked found bed control switch not working Found control pcb defective Repaired pcb with genuine components and control working good. Pcb defective due to technical issues			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer Jince cherian ATCO Technology Co. Signature  Date 22/06/25		Authorized Person  Signature Date مسجل طب طبيعي وتأهيل مستشفى البصرة 222100196	
Customer / User Signature Date			