

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



Log no - 2161

شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجاهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٢٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61986

Customer	Alorf Hospital	Equipment	OT Table.
		Model No.	V750
		Serial No.	NV111020091
Authorized Person		Manufacturer	Nuvo
Customer Complaint		Location	O.R
Table is not stable		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		

Remarks	Service call attended.
	We inspect the unit and found that the
	screws are loose. we tighten the screws
	and checked the unit.
	checked and confirmed that the unit is
	working properly.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer	Authorized Person	Customer / User
M. Alon / M. Alon		
ATCO Technology Co.		
Signature	Signature	Signature
Date 21/08/2024	Date	Date