

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO . 336218



شركة اتكو التكنولوجيه
تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)
فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)
ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت
Email: sales@atco-kuwait.com
www.atco-kuwait.com
س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 56694

Customer	Alor Hospital.			Equipment	ENT Workstation		
				Model No.	FUTURENT III		
				Serial No.	111632045		
Authorized Person				Manufacturer	ENTERMED		
Customer Complaint				Location	ENT		
water injector screw removed from the connector				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation		☐ Warranty	
				☒ Repair		☐ On Charge	
				☐ Maintenance		☐ Contract	
				☐ Upgrade		☐	
				Time Spent			
Remarks Service call attended. while we inspect the unit, the water irrigation connector is removed from the hand piece. we re connect the @ connector. checked and confirmed that there is no leakage. The unit is working good.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer Nixon ATCO Technology Co. Signature Date 14/01/2024				Authorized Person Signature Date		Customer / User Signature 14.1.24 Date	