

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة اتكو التكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 59406

Customer	Alord Hospital	Equipment	ECC
		Model No.	CARDIPIA 800
		Serial No.	C80071048 LHM 58125R
Authorized Person		Manufacturer	TRISMED
Customer Complaint		Location	Q.B
		Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☒ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Optics Checks	☐ Maintenance
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade
☐ Safety Checks	☐ Operational Error	☐
☐ Electrical / Electronics	☐ Mechanical Checks	
☐ Accessories Check	☐	
	Time Spent	

Remarks	Service call attended. we inspect the unit and checked the ECC using Stimulator. we found that the ECC cable is faulty. Replacement of ECC cable, checked and confirmed that the unit is working properly.
---------	--

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Nixon ATCO Technology Co. Signature Date 06/04/2021	Authorized Person Signature Date 16/4/2021	Customer / User Signature Date
--	--	--------------------------------------