

Customer <u>Alorif Hospital</u>		Equipment <u>CSSD Washer & Autoclave</u>	
		Model No. <u>water filter.</u>	
		Serial No. <u>230504</u>	
Authorized Person		Manufacturer	
Customer Complaint		Location <u>CSSD</u>	
		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐ _____		☐ Installation ☐ Warranty	
		☐ Repair ☐ On Charge	
		☒ Maintenance ☐ Contract	
		☐ Upgrade ☐ _____	
		Time Spent	
Remarks <u>water filter replacement.</u> <u>Replaced the six water filter with</u> <u>new one -</u> <u>checked and confirmed that the</u> <u>unit is working properly.</u>			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer <u>Nixon/many</u> ATCO Technology Co. Signature <u>[Signature]</u> Date <u>14/01/2024</u>		Authorized Person <u>[Signature]</u> <u>20/1/24</u> Signature Date	
		Customer / User Signature Date	