

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب. ٢٣٥٥ الجهره, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61982

Customer	Alorf Hospital.			Equipment	DVT - Compressed limb therapy		
				Model No.	DVT 7700		
				Serial No.	P770B140K0004		
Authorized Person				Manufacturer	Wanji		
Customer Complaint	cuff limb connector broken.			Location	5A WAP		
				Contract / Order No.			
Services / Check Carried out	☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check			☐ Installation ☐ Repair ☐ Maintenance ☐ Upgrade	☐ Warranty ☐ On Charge ☐ Contract ☐		
	☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐			Time Spent			
Remarks	Service call attended. we inspect the unit and found that the limb cuff connector in machine side broken. we replaced the connector from the spare machine in store. checked and confirmed that the unit is working properly.						
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer	Authorized Person			Customer / User			
Signature	Signature			Signature			
Date	Date			Date			