



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 02/01/25

Requested by Prasanth

Department C.S.S.D Location CSSD Basement

SR. No.	Machine	Model	Complaint
<u>230504</u>	<u>washer disinfection</u>	<u>aniclean P2 II</u>	<u>water spray not working</u>

Check within Week

ASAP

Engineer Comments

End Comments

Prasanth
Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرات, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 64286

Customer	Al Ory Hospital	Equipment	Washer
	CSSD	Model No.	
		Serial No.	Washer 1 & 2
Authorized Person		Manufacturer	MMM
Customer Complaint		Location	CSSD
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
		Time Spent	

Remarks

Modification of washer disinfectant supply line.
Water supply connected to water softener system

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

ATCO Service Team

ATCO Technology Co.

Signature

Date

Authorized Person

Signature

Date

Customer / User

Signature

Date