

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء ، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٢٣٦٢٨

SERVICE / INSTALLATION REPORT

No.: 63214

Customer	Alot Hospital	Equipment	Defibrillator
		Model No.	Cardio Aid 360 B.
		Serial No.	212060645
Authorized Person		Manufacturer	Innomed
Customer Complaint		Location	Cardio 4th Floor.
		Contract / Order No.	
		☐ Installation	☐ Warranty
		☒ Repair	☐ On Charge
		☐ Maintenance	☐ Contract
		☐ Upgrade	☐ <u>EM</u>
		Time Spent	
Services / Check Carried out			
☐ Inspection / Cleaning	☐ Optics Checks		
☐ Calibration / Performance	☐ Pneumatic Checks		
☐ Interface / Communication	☐ Hydraulic Checks		
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	Unit under Battery Formatting. To be checked again if same problem Battery need to be replaced.		

No.	Spare Parts Used	Part No.	Qty.

Service Engineer ATCO Technology Co. Signature Date 3/4/24	Authorized Person Signature Date	Customer / User Signature Date 22022100348
---	--	--